

Product:**Code Number:****Capto™ Q**

17531604

Lot No: 10333558

Test/Characteristic:	Limit:	Result:
Ionic capacity mmol Cl ⁻ / mL resin	0.16 - 0.22	0.18
Breakthrough capacity, Q_{B,10%} mg BSA / mL resin	min. 100	128
Microbial contamination Colony Forming Units / mL suspension	max. 20	0
Endotoxin activity EU/mL	less than 5.0	<0.25

**Manufactured in compliance with our ISO 9001 certified quality management system.
No animal derived material is used in raw materials or in the manufacture of this product.**

Approval Date (Year-Month-Day): 2022-12-19**Expiry Date (Year-Month): 2027-11****Manufacturing Date (Year-Month): 2022-11**